



Internationally indexed journal

Indexed in Chemical Abstract Services (USA), Index copernicus, Ulrichs Directory of Periodicals, Google scholar, CABI ,DOAJ , PSOAR, EBSCO , Open J gate , Proquest , SCOPUS , EMBASE ,etc.



Rapid and Easy Publishing

The "International Journal of Pharma and Bio Sciences" (IJPBS) is an international journal in English published quarterly. The aim of IJPBS is to publish peer reviewed research and review articles rapidly without delay in the developing field of pharmaceutical and biological sciences



Pharmaceutical Sciences

- Pharmaceutics
- Novel drug delivery system
- Nanotechnology
- Pharmacology
- Pharmacognosy
- Analytical chemistry
- Pharmacy practice
- Pharmacogenomics



Biological Sciences

- Polymer sciences
- Biomaterial sciences
- Medicinal chemistry
- Natural chemistry
- Biotechnology
- Pharmacoinformatics
- Biopharmaceutics
- Biochemistry
- Biotechnology
- Bioinformatics
- Cell biology
- Microbiology
- Molecular biology
- Neurobiology
- Cytology
- Pathology
- Immunobiology

**Indexed in Elsevier Bibliographic Database
(Scopus and EMBASE)**

SCImago Journal Rank 0.288

Impact factor 2.958*

Chemical Abstracts
Service (www.cas.org)



A division of the American Chemical Society

CODEN IJPBJ2



Elsevier Bibliographic databases (Scopus & Embase)

SNIP value – 0.77

SJR - 0.288

IPP - 0.479

SNIP – Source normalised impact per paper

SJR – SCImago Journal rank

IPP – Impact per publication

Source – www.journalmetrics.com

(Powered by scopus (ELSEVIER))



LUND
UNIVERSITY



JACKSONVILLE STATE UNIVERSITY

Jacksonville State University
Houston Cole Library
USA (Alabama)



UNIVERSITY OF
OXFORD

Oxford, United Kingdom

INDEX COPERNICUS
INTERNATIONAL

*And indexed/catalogued in
many more university*



*Instruction to Authors visit www.ijpbs.net

For any Queries, visit "contact" of www.ijpbs.net



PERFORMANCE EVALUATION OF SPECKLE REDUCTION FILTERS FOR OPTICAL COHERENCE TOMOGRAPHY IMAGES

VENKATRAMAN.K*¹ AND M.SUMATHI²

¹Research Scholar, Sathyabama University, Chennai - 119, Tamilnadu, India.

²Professor, Electronics and Control, Sathyabama University, Chennai-119, Tamilnadu, India.

ABSTRACT

Medical Images play a vital role in the diagnosis of various diseases. Hence high resolution and reduced noise becomes vital while representing these images. The Medical imaging modalities using Ultrasound produces images which are affected by speckle noises which are multiplicative in nature, caused due to interference of the backscattered image. One such modality used in ophthalmology is the Optical Coherence Tomography. Various filtering techniques have been proposed to reduce these speckle noises by pertaining the required image information. This paper aims at evaluating various such proposed filters. OCT image acquired using Karl Zeiss was used for this purpose of evaluation. Various filters were tested for speckle affected OCT image and the performances of these filters were evaluated in terms of Mean Square Error, Signal to Noise Ratio, Peak Signal to Noise Ratio, and visual inspection. The system was designed using MATLAB, Intel core 2 duo processor and was tested with images obtained from a medical research center. Results show that homomorphic wiener filters show a better performance followed by Gaussian filters. The results of the tested system for OCT Images can also be generalized for any other ultrasound based modality images.

KEYWORDS: *Interference, Speckles, Image Smoothing, Mean Square Error, Signal to noise ratio*

*Corresponding author



VENKATRAMAN.K

Research Scholar, Sathyabama University, Chennai - 119,
Tamilnadu, India.

I. INTRODUCTION

Medical Images are more prone to noise in the stage of acquisition. It is essential to remove these noises by pertaining all the required informations, so that the image quality and informations are least affected. Optical Coherence Tomography is one of the widely used procedures, which serves as a vital tool of diagnosis in the field of ophthalmology, except for a disadvantage that these images are corrupted by speckle noises. OCT cross sectionally measures the optical reflections from the internal structures of the biological tissues¹. For a reliable diagnosis of various retinal disorders, it is essential to obtain a high resolution Image. The measurements of Intra Ocular Pressure do not necessarily predict the progression of various diseases like glaucoma². Loss of vision and Optic Nerve Head cupping are being detected only after 50% vision loss³. It is also necessary to properly image the retinal nerve fiber layer (RNFL) which is comparatively a layer of higher scattering than any other retinal structures. The backscattered light from RNFL strongly depends upon the incident angle of light^{4,5}. These reflecting and backscattering rays raises some practical limitations on OCT. Independent of the texture of the surface, these backscattered and reflected light waves tend to interfere, thereby causing dark and bright spots. These noises are referred to as speckle noises, which are multiplicative in nature. This speckle formation is influenced by various parameters of the light source and detectors. Speckles are commonly formed due to multiple backscattering of beam and random delays in forward propagating and returning beams. When a large number of polarized quasi monochromatic waves with random phase combine, a fully developed speckle pattern is formed. Note that if all the backscattered are made to constructively interfere, it could be seen that the noise would disappear and contrast would be greatly enhanced. Speckles could be classified generally as signal carrying and signal degrading speckles, based on their origin. Speckles caused due to single backscattering are generally more informative and signal carrying in nature. It is necessary to

preserve these informations howmuch ever possible while preprocessing the images. These types of speckles could also be differentiated by the spot sizes of correlation. Various filters used in speckle noise reduction are evaluated in terms of mean square error, signal to noise ratio, as well as improvisation in visual quality. The research work explains various filters used in reduction of speckle noises and the experimental analysis of these filters and their performance.

II. SPECKLE NOISE REDUCTION

Various methodologies proposed for reduction of speckle noises include, median filters⁶⁻⁸, homomorphic wiener filter⁹, multiresolution wavelet analysis¹⁰, adaptive smoothing¹¹, frost filter¹²⁻¹⁴, diffusion filter¹⁵⁻¹⁹. It could be understood that the appropriate method of reducing the speckle noises is the one which enhances the image by preserving the edges and lines in the image^{7,20,21}.

Mean Filters

Mean filters are simple smoothing filters which reduces the amount of variation in intensity between the neighbor pixels of the image. A window of standard sizes is generated and spatial filtering is performed using this window on the image. This filter computes the sum of all pixels in the slided filter window and then divides the sum by the number of pixels in the filter window²²⁻²⁴. The algorithm is so simple. The window is placed over the image. Average is taken (Sum up the elements and divide the sum by the number of elements).

New pixel value = Sum (W*image pixel value)/ Number of pixels (2.1)

where *W* is the weight factor, which is the ratio of standard deviation and mean. It could be understood that when a mean filter is used, a single pixel with a very unrepresentative value can significantly affect the mean value of all the pixels in its neighborhood. The major disadvantage is that when the filter neighborhood straddles an edge, the filter will

interpolate new values for pixels on the edge and so will blur that edge. This may be a problem if sharp edges are required in the output.

Frost Filters

Designing of frost filters are dependent on factors that Speckle noise is a multiplicative noise, i.e. it is in direct proportion to the local grey level in any area. Also the signal and the noise are statistically not dependent on one another. It is exponentially weighted averaging filter which is dependent on the variation coefficient. The coefficient of variation is given by the ratio of local standard deviation to the local mean of the noise corrupted image. These are weighted filters whose weight factors decrease with distance from the pixel of interest.

$$\text{New pixel value} = \text{Sum} (W * e^{-Wt}) \quad (2.2)$$

Adaptive smoothing filters

These are conventional filters of smoothing except that smoothing is done pixel wise iteratively, based on the input image and intensity of the noise. The filter coefficients adapts to the local changes in image statistics for n iterations specified by the user. For multiplicative noise, the adaptive noise smoothing filter is a systematic derivation of Lee's algorithm with some extensions that allow different estimators for the local image variance²⁵. The filter entirely depends on the image statistics and no other priori informations on the original image are required.

Gaussian Filters

It is similar to any other smoothing filter except that the extent of smoothing/ noise removal is dependent on the standard deviation of the Gaussian. It is a mean filter which uses a different kernel that gives the shape of a Gaussian ('bell-shaped') hump. It could be understood that the weights of the spatial filters varies based on the standard deviation. The surrounding pixels with lesser deviation and closer as neighbour is given with a higher weightage, whereas, pixels with higher

deviation and distant from the center pixel is given with a lesser weightage

Bilateral Filters

A bilateral filter is a non-linear filter which could be understood as an advancement of Gaussian filter in which the weighting of the filter coefficients is with the corresponding relative pixel intensities. It aims at smoothing the image by preserving the edges. Each pixel is replaced by weighted average of intensity values of the neighbourhood pixels. Pixels that are very different in intensity from the central pixel are weighted less even if they are close neighbours to the central pixels. This weight can be based on a Gaussian distribution. The weights depend on the euclidean distance of pixels, and the radiometric differences. Various artifacts included in image due to bilateral filtering include Staircase effect and Gradient reversal (introduction of false edges).

$$\text{New pixel value} = \text{Avg} (W * \text{Image pixel value}) \quad (2.3)$$

where, *W* is the weight which is varied for every surrounding pixels

Anisotropic Diffusion Filters

Diffusion could be understood as balancing the changes in concentration of an image. Anisotropic diffusion is a generalization of this diffusion process, which produces a family of parameterized images, but each resulting image is a combination between the original image and a filter that depends on the local content of the original image. The rate to which the image should be diffused is dependent on the divergence of the surrounding pixels with respect to the center pixel. These filters are more commonly used in segmentation and edge detection, as they are comparatively good in preserving the edges even while smoothing the other regions of the image.

Homomorphic Wiener and Wavelet Filters

Homomorphic filters are commonly used in removing multiplicative noises. It normalizes the brightness throughout the image and enhances

the contrast. Not restricting to preprocessing, these filters can also be used in enhancing the image. Illumination and reflectance components of an image are separated and then wiener and wavelet filters are applied to these components separately, after which it is recombined. The multiplicative noises are made additive by taking logarithm. These noises which are made additive are then removed linearly using wavelet and wiener filters. The basic idea is to convert the multiplicative noise into additive component and then remove them.

III. EXPERIMENTAL ANALYSIS AND DISCUSSION

Various filters as explained in section 2 were designed and evaluated for a speckle noise affected Optical Coherence Tomography image. Fig. 1 shows an OCT image with speckle noises and its monochrome image. The performances of the filters are evaluated in terms of the mean squared error (MSE), Signal to noise Ratio (SNR) and Peak Signal to Noise Ratio (PSNR).




Figure 1
Speckle noise affected OCT Image.

Mean Squared Error is the average value of the squared values of the error between the original image and the estimated image. Signal to noise ratio gives the ratio of the concentration of the signal/required data with that of the noise. Peak Signal to Noise ratio gives the ratio of peak power of the signal with that of the noise. These parameters are expressed in equations 3.1, 3.2, 3.3.

$$MSE = \text{Mean}(\text{error}^2) \quad (3.1)$$

$$SNR = \text{Signal power}/\text{Noise power} \quad (3.2)$$

$$PSNR = \text{Peak Signal power}/\text{Peak Noise power} \quad (3.3)$$

For an image to be with more desirable qualities, the SNR and PSNR are expected to be greater, which means that the intensity of noise must be much lesser. The outputs of various filters with various parameters are shown in fig. 2 and fig. 3.




Figure 2

Output of various speckle reduction filters. (a) Mean filter 3*3, (b) Mean filter 5*5, (c) Mean filter 7*7, (d) Frost filter, (e) Adaptive smoothing filters, (f) Gaussian Filters 3*3, (g) Gaussian Filters 5*5, (h) Gaussian Filters 7*7, (i) Bilateral filters, (j) Anisotropic diffusion filters, (k) Homomorphic Wiener filters 3*3, (l) Homomorphic Wiener filters 5*5.




Figure 3
Output of various band eliminated homomorphic wavelet (db2) filters

Filter	MSE	SNR	PSNR
Mean (3*3)	0.0035	72.2718	72.6871
Mean (5*5)	0.0078	68.7923	69.2076
Mean (7*7)	0.0103	67.5839	67.9992
Adaptive smoothing	0.0106	66.9773	67.8866
Anisotropic Diffusion	0.0076	68.8824	69.2977
Frost	0.0028	73.2101	73.6254
Gaussian (3*3)	0.00005	81.048	81.4633
Gaussian (5*5)	0.00005	81.0183	81.4337
Gaussian (7*7)	0.00005	81.0183	81.4337
Homomorphic wiener	$1.2 \cdot 10^{-7}$	97	97
Homomorphic wavelet - LH	2.1599	44.3712	44.7865
Homomorphic wavelet - HL	2.1613	44.3684	44.7837
Homomorphic wavelet - HL – LH	2.1723	44.3463	44.7616
Homomorphic wavelet - HL – HH	2.1433	44.4046	44.8199
Homomorphic wavelet - LH – HH	2.1557	44.3796	44.7949
Homomorphic wavelet - HH	2.1544	44.3822	44.7975

Table 1
Performance metrics of filters in speckle noise reduction

The SNR and PSNR parameters of the various filters used are represented in the fig. 4 and fig. 5.




Figure 4
SNR Values of various filters




Figure 5
PSNR Values of various filters

From the results obtained, it could be seen that homomorphic filters are more effective in removing the speckle noises for the OCT images. The homomorphic wiener filter seems

to show a better performance not only in analysis of the statistical parameters but also from the visual inspection.

IV. CONCLUSION

It could be seen that among the various filters evaluated, both statistically as well as visually. Most of the filters fail to remove the speckle noises in medical image, since the details about noise variance may not be identified. Most of the filters just average the data instead of removing the speckles. The computational results show that Homomorphic wiener filter is comparatively better, the reason being that the multiplicative noises are made additive and hence noise removal is much effective comparatively. Gaussian filter and frost filters

also shows appreciable results following the homomorphic wiener filter. It could also be seen that irrespective of the size of the kernel, mean filter showed a poor performance. This is justifiable as the mean filters do not take more statistical properties into consideration. Also it could be seen that Homomorphic wavelet filters show poor performance than any other filters. The results are evaluated with OCT images, but are applicable for any imaging modality with speckle noises. The results are expected to be similar for other medical images including Ultrasound.

V. REFERENCES

1. David Huang, Eric. A. Swanson, Charles. P. Lin, Joel. S. Schuman, Willim. G. Stinson, Warren Chang, Michael. R. Hee, Thomas Flotte, Kenton Gregory, Carmen. A. Puliafito, James. G. Fujimoto. Optical Coherence Tomography. *J Stor. Science*, 254(5035): 1178-1181, (1991).
2. A.Sommer. Intraocular Pressure and Galucoma. *Am. J. Ophthalmol.*, 107: 186-188, (1989).
3. H. A. Quigley, E. M. Addicks, *Arch. Ophthalmol.* 100: 807 (1982).
4. R. W. Knighton, S.G. Jacobson, C. M. Kemp, *Inv. Ophthalmol. Vis. Sci.* 30: 2393, (1989).
5. J. W. Goodman, *Statistical Optics*, Wiley, New York, pp. 124-127; 1985.
6. R. Bernstein, Adaptive nonlinear filters for simultaneous removal of different kinds of noises in images, *IEEE Trans. On Circuits and Systems*, 34: 1275-1291 (1987).
7. Achim Bezerianos A. and Tsakalides. P, Novel Bayesian Multiscale for Speckle Removal in Medical Ultrasound Images. *IEEE Transactions on Medical Imaging Journal*, 20(8): 772-783, (2001).
8. J. G. Abbott and F. L. Thurstone, Acousticspeckle: Theory and experimental analysis, *Ultrason. Imag.*, 1: 303-324, (1979).
9. G. Franceschetti, V. Pascazio, and G. Schirinzi, Iterative homomorphic technique for speckle reduction in synthetic-aperture radar imaging, *J. Opt. Soc. Am.* 12: 686-694, (1995).
10. P. Moulin, A Wavelet regularization method for diffuse radar-target imaging and speckle-noise reduction, *J. Math. Imag. Vision* 3: 123-134, (1993).
11. D. T. Kuan, A. A. Sawchuk, T. C. Strand, and P. Chavel, Adaptive noise smoothing for images with signal-dependent noise, *IEEE Trans. Patt. Anal. Mach. Intell.* 7:165-177 (1985).
12. Raman Maini, Himanshu Aggarwal, Performance Evaluation Of Various Speckle Noise Reduction Filters On Medical Images. *International Journal of Recent Trends in Engineering*, 2(4): 22-25, (2009).
13. J.S. Lee, Refined filtering of image noise using local statistics, *Computer Vision, Graphics, and Image Processing*.
14. V. S. Frost, J. A. Stiles, K. S. Shanmugam, J.C. Holtzman, A model for radar image & its application To Adaptive digital filtering for multiplicative noise, *IEEE Transaction on pattern analysis and machine intelligence*, 4: 175, (1982).
15. Abrishami Moghaddam. H, ValadanZouj. M.J, Dehghani. M, Bayesian-based speckle reduction using wavelet transforms, *Conference on applications of Digital Image Processing in 49th SPIE annual meeting*, (2004).

16. K. Krissian, G. Malandain, and N. Ayache, Directional Anisotropic Diffusion applied to Segmentation of Vessels in 3D images, Research Report No.3064, INRIA Sophia Antipolis, (1996).
17. N. Ayache. "Volume Image processing, Results and Research Challenges" Research report, No. 2050, INRIA, (1993).
18. A. Herment, J.P. Guolielmi, P. Dumeé, P. Peronneau, and P. Delouche., Limitations of ultrasound imaging and imaging restoration., *Ultrasonics*, (1987)
19. J.S. Lee., Digital Image enhancement and noise filtering by use of local statistics., *IEEE Transactions on pattern analysis and machine intelligence*, 2(2):165-168, (1980).
20. R.C. Gonzalez and R.E. Woods: 'Digital Image Processing', Addison-Wesley Publishing Company, 2002.
21. Image Processing Fundamentals – Statistics, "Signal to Noise Ratio", 2001.
22. Kenneth R. Castleman, *Digital Image Processing*. Prentice Hall, 1996.
23. Anil K. Jain, *Fundamentals of Digital Image Processing*. Prentice Hall, 1989.
24. William K. Pratt, *Digital Image Processing*. Wiley, 1991.
25. Kuan DT, Sawchuk AA, Strand TC, Chavel P., Adaptive noise smoothing filter for images with signal-dependent noise, *IEEE Trans Pattern Anal Mach Intell.*, 7(2): 165-177, (1985).